

To whom it may concern

Radiometer Medical ApS
Åkandevvej 21
2700 Brønshøj
Denmark
Phone: +45 38 27 38 27
Fax: +45 38 27 27 27
CVR no. 27509185
www.radiometer.com

April 12, 2022

Spring4Shell Security Vulnerability

Dear valued customer

A vulnerability named "Spring4Shell" (CVE-2022-22963 and CVE-2022-22965), has been identified in the Spring framework used for web system development. This vulnerability can when exploited allow an attacker to execute remote code.

Radiometer Medical has evaluated whether these vulnerabilities have an impact on our products:

- ABL80
- ABL800
- ABL9
- ABL90
- AQT90
- TCM5
- AQUIRE

The framework is not used in any of our products and have therefore no impact on any of the above devices.

This message constitutes an information service and is provided without guarantees and warranties of any kind, expressed or implied. Radiometer excludes any liability for actions or omissions performed based on this information.

Kind regards



Christian Knudsen
Technical Product manager